

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: LH320166G

Date Tested: 20-Jun-17

Instrument Number: 8470842

Mac Address: 00:0A:8D:81:41:3A

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.789	382	6.018	882	9.240	1382
Channel 2	2.796	383	6.024	884	9.248	1384
Channel 3	2.792	382	6.025	883	9.248	1383
Channel 4	2.802	384	6.030	885	9.255	1384
Channel 5	2.798	383	6.020	883	9.256	1385
Channel 6	2.797	383	6.027	884	9.258	1385

Allowable difference in temp reading is +or- 5 Deg. F.

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2097001

Calibration Date: 13-Mar-17

Precision digital multimeter SN: 2142012

Calibration Date: 13-Mar-17

ProHeat 35 power supply used for signal output SN: ME040114G


Calibration date: 20-Jun-17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician:



Calibration due date: 20-Jun-18

Signature