

## Calibration Certificate

Unit Serial Number: LJ330096G

Date Tested: 5/1/14

Instrument Number: 62074

Mac Address: 00:0A:8D:F2:7A

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.782	382	6.013	882	9.235	1382
Channel 2	2.788	382	6.011	882	9.236	1382
Channel 3	2.788	382	6.012	882	9.238	1382
Channel 4	2.787	382	6.011	882	9.236	1382
Channel 5	2.784	382	6.009	882	9.235	1382
Channel 6	2.785	382	6.010	882	9.234	1382

**Allowable difference in temp reading is +or- 5 Deg. F.**

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2083066

Calibration Date: 7/12/13

Precision digital multimeter SN: 2141018

Calibration Date: 7/10/13

ProHeat 35 power supply used for signal output SN: MD150006G

Calibration date: 5/1/14

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: DON WAGNER

Calibration due date: 5/1/2015

Signature