

## Calibration Certificate

Unit Serial Number: MA120021G

Date Tested: 10/17/12

Instrument Number: 95032

Mac Address: 00:0A:8D:01:73:38

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.787	382	5.688	832	9.240	1383
Channel 2	2.788	382	5.690	832	9.239	1383
Channel 3	2.787	382	5.689	832	9.240	1383
Channel 4	2.787	382	5.690	833	9.240	1382
Channel 5	2.788	382	5.690	833	9.239	1382
Channel 6	2.788	382	5.690	832	9.239	1383

**Allowable difference in temp reading is +or- 5 Deg. F.**

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2083066

Calibration Date: 6/25/12

Precision digital multimeter SN: 214018

Calibration Date: 8/15/12

ProHeat 35 power supply used for signal output SN: LJ430396

Calibration date: 6/27/12

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Don Wagner

Calibration due date: 10/17/13

Signature