

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MC350423G
Instrument Number: 8463720
IP Address: 192.168.111.222

Date Tested: MARCH 2/17
Mac Address: 00:0A:8D:81:25:68

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.787	382	6.013	882	9.239	1382
Channel 2	2.789	382	6.012	882	9.237	1382
Channel 3	2.791	382	6.016	882	9.238	1382
Channel 4	2.788	382	6.015	882	9.239	1382
Channel 5	2.785	382	6.015	882	9.240	1382
Channel 6	2.790	382	6.016	882	9.239	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 9/21/16

Precision digital multimeter SN: 2142011

Calibration Date: 9/21/16

ProHeat 35 power supply used for signal output SN: MD130155 Calibration date: 2/16/17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.
Calibration due date:

Technician: D. Burger
Signature