

Red-D-Arc Welderentals

Calibration Certificate

Unit Serial Number: MC350425G Date Tested: MAR. 1/17
 Instrument Number: 8459777 Mac Address: 00:0A:8D:81:16:01
 IP Address: 192.168.111.222

| Primary Standard | | | | | | |
|------------------|----------------|------------------|----------------|------------------|----------------|------------------|
| | 382 (F) | | 882 (F) | | 1382 (F) | |
| | Signal Voltage | Recorder display | Signal Voltage | Recorder display | Signal Voltage | Recorder display |
| Channel 1 | 2.786 | 382 | 6.013 | 882 | 9.238 | 1382 |
| Channel 2 | 2.788 | 382 | 6.015 | 882 | 9.239 | 1382 |
| Channel 3 | 2.788 | 382 | 6.018 | 882 | 9.238 | 1382 |
| Channel 4 | 2.786 | 382 | 6.013 | 882 | 9.238 | 1382 |
| Channel 5 | 2.786 | 382 | 6.010 | 882 | 9.238 | 1382 |
| Channel 6 | 2.787 | 382 | 6.015 | 882 | 9.238 | 1382 |

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 9/21/16

Precision digital multimeter SN: 2142011

Calibration Date: 9/21/16

ProHeat 35 power supply used for signal output SN: MD130155

Calibration date: 2/16/17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date:

Technician: D. Burger

Signature