



Miller Electric Mfg. Co.
1635 West Spencer Street
P.O. Box 1079
Appleton, WI 54912-1076

CERTIFICATE OF CALIBRATION VERIFICATION

PROHEAT Serial Number: MG040199G

TCI Serial Number: J1505380040

Certified by: WP

Verification Date: Jan 20, 2016

Verification Due: Jan 19, 2017

Miller Electric Mfg. Co. does hereby certify the above instrument was calibrated against standards maintained by Miller Electric Mfg. Co. and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Instrumentation Used:

Fluke Model No. 714 Thermocouple Calibrator

Agilent Multimeter Model 34401A

Serial Number: L04275

Serial Number: L04222